



## Certificate of Analysis

<b>Client:</b>	Hawkes Bay Regional Council	<b>Lab No:</b>	1971361	SPV1
<b>Contact:</b>	V Lyon C/- Hawkes Bay Regional Council Private Bag 6006 Napier 4142	<b>Date Received:</b>	28-Apr-2018	
		<b>Date Reported:</b>	11-May-2018	
		<b>Quote No:</b>	81040	
		<b>Order No:</b>	RM193	
		<b>Client Reference:</b>	Whangawehi (one only)	
		<b>Add. Client Ref:</b>	312-302	
		<b>Submitted By:</b>	V Lyon	

### Sample Type: Aqueous

<b>Sample Name:</b>	63192 - Whangawehi @ George Ormonds - 3301 27-Apr-2018 1:30 pm				
<b>Lab Number:</b>	1971361.1				

### Faecal Coliforms and E. coli profile

Faecal Coliforms	cfu / 100mL	53	-	-	-	-
Escherichia coli	cfu / 100mL	48	-	-	-	-

### HBRC Standard River

Volatile Suspended Solids	g/m <sup>3</sup>	0.5	-	-	-	-
Total Suspended Solids	g/m <sup>3</sup>	1.8	-	-	-	-
Total Nitrogen	g/m <sup>3</sup>	0.31	-	-	-	-
Total Ammoniacal-N	g/m <sup>3</sup>	< 0.005	-	-	-	-
Nitrite-N	g/m <sup>3</sup>	0.0019	-	-	-	-
Nitrate-N	g/m <sup>3</sup>	0.024	-	-	-	-
Nitrate-N + Nitrite-N	g/m <sup>3</sup>	0.026	-	-	-	-
Total Kjeldahl Nitrogen (TKN)	g/m <sup>3</sup>	0.29	-	-	-	-
Dissolved Reactive Phosphorus	g/m <sup>3</sup>	0.037	-	-	-	-
Total Phosphorus	g/m <sup>3</sup>	0.058	-	-	-	-

### Analyst's Comments

Please interpret this result with caution as the sample was > 8 °C on receipt at the lab. The sample temperature is recommended by APHA to be less than 8 °C on receipt at the laboratory (but not frozen). However, it is acknowledged that samples that are transported quickly to the laboratory after sampling, may not have been cooled to this temperature.

## Summary of Methods

The following table(s) gives a brief description of the methods used to conduct the analyses for this job. The detection limits given below are those attainable in a relatively clean matrix. Detection limits may be higher for individual samples should insufficient sample be available, or if the matrix requires that dilutions be performed during analysis.

### Sample Type: Aqueous

Test	Method Description	Default Detection Limit	Sample No
Individual Tests			
Filtration, Unpreserved	Sample filtration through 0.45µm membrane filter.	-	1
Volatile Suspended Solids	Filtration (GF/C, 1.2 µm). Ashing 550°C, 30 min. Gravimetric. APHA 2540 E (modified) 22 <sup>nd</sup> ed. 2012.	0.5 g/m <sup>3</sup>	1
Total Suspended Solids	Filtration of a 2L sample using Whatman 934 AH, Advantec GC-50 or equivalent filters (nominal pore size 1.2 - 1.5µm), gravimetric determination. APHA 2540 D (modified) 22 <sup>nd</sup> ed. 2012.	0.5 g/m <sup>3</sup>	1



Sample Type: Aqueous			
Test	Method Description	Default Detection Limit	Sample No
Total Nitrogen	Calculation: TKN + Nitrate-N + Nitrite-N. Please note: The Default Detection Limit of 0.05 g/m <sup>3</sup> is only attainable when the TKN has been determined using a trace method utilising duplicate analyses. In cases where the Detection Limit for TKN is 0.10 g/m <sup>3</sup> , the Default Detection Limit for Total Nitrogen will be 0.11 g/m <sup>3</sup> .	0.05 g/m <sup>3</sup>	1
Total Ammoniacal-N Trace	Phenol/hypochlorite colorimetry. Flow injection analyser. (NH <sub>4</sub> -N = NH <sub>4</sub> <sup>+</sup> -N + NH <sub>3</sub> -N). APHA 4500-NH <sub>3</sub> H 22 <sup>nd</sup> ed. 2012.	0.005 g/m <sup>3</sup>	1
Nitrite-N Trace	Automated Azo dye colorimetry, Flow injection analyser. APHA 4500-NO <sub>2</sub> -I 22 <sup>nd</sup> ed. 2012 (modified).	0.0010 g/m <sup>3</sup>	1
Nitrate-N	Calculation: (Nitrate-N + Nitrite-N) - NO <sub>2</sub> N. In-House.	0.0010 g/m <sup>3</sup>	1
Nitrate-N + Nitrite-N Trace	Total oxidised nitrogen. Automated cadmium reduction, flow injection analyser. APHA 4500-NO <sub>3</sub> -I 22 <sup>nd</sup> ed. 2012 (modified).	0.0010 g/m <sup>3</sup>	1
Total Kjeldahl Nitrogen (TKN)	Total Kjeldahl digestion, phenol/hypochlorite colorimetry. Discrete Analyser. APHA 4500-Norg D. (modified) 4500 NH <sub>3</sub> F (modified) 22 <sup>nd</sup> ed. 2012.	0.10 g/m <sup>3</sup>	1
Dissolved Reactive Phosphorus	Filtered sample. Molybdenum blue colourimetry. Flow injection analyser. APHA 4500-P G (modified). 22 <sup>nd</sup> ed. 2012.	0.004 g/m <sup>3</sup>	1
Total Phosphorus	Total phosphorus digestion, ascorbic acid colorimetry. Discrete Analyser. APHA 4500-P B & E (modified from manual analysis) 22 <sup>nd</sup> ed. 2012. Also modified to include the use of a reductant to eliminate interference from arsenic present in the sample. NAWASCO, Water & soil Miscellaneous Publication No. 38, 1982.	0.004 g/m <sup>3</sup>	1
HBRC Standard River		0.0010 - 0.5 g/m <sup>3</sup>	1
Faecal Coliforms and E. coli profile			
Faecal Coliforms	Membrane Filtration, Count on mFC agar, Incubated at 44.5°C for 22 hours, Confirmation. APHA 9222 D, 22 <sup>nd</sup> ed. 2012.	1 cfu / 100mL	1
Escherichia coli	Membrane filtration, Count on mFC agar, Incubated at 44.5°C for 22 hours, MUG Confirmation. APHA 9222 G, 22 <sup>nd</sup> ed. 2012.	1 cfu / 100mL	1

These samples were collected by yourselves (or your agent) and analysed as received at the laboratory.

Samples are held at the laboratory after reporting for a length of time depending on the preservation used and the stability of the analytes being tested. Once the storage period is completed the samples are discarded unless otherwise advised by the client.

This certificate of analysis must not be reproduced, except in full, without the written consent of the signatory.

Kim Harrison MSc  
Client Services Manager - Environmental